

# Clinical Chemistry TCPLA\_CBC\_003

## Purpose

Clinical chemistry determines biochemical parameters in plasma including enzymatic activity, specific substrates and electrolytes.

Ontological description: MP:0001545 – blood physiology abnormalities.

## Experimental Design

- **Minimum number of animals** : 7M + 7F
- **Age at test**: Week 59
- **Sex**: We would expect the results of this test to show sexual dimorphism

## Equipment

1. Clinical chemistry analyser
2. Vortex
3. Refrigerated centrifuge
4. Eppendorf tubes
5. Pipettes (200-1000 ul)

## Procedure

Set up the clinical chemistry analyser and perform QC analyses of the control reagents in accordance with the equipment guidelines.

### Sample collection and preparation:

- a. Collect the appropriate volume of blood required (160-200l of plasma), for the clinical chemistry analyser being used for assessment, preferably using a tube with an anticoagulant with the relevant blood collection procedure (see IMPC protocol Blood collection by retro-orbital puncture). Time of day for collection is in the morning, starting no earlier than 07:30.
- b. Keep whole blood samples on ice until centrifugation. Centrifuge for 10 minutes at 5000 x g in a refrigerated centrifuge set at 8°C. If plasma samples cannot be analysed immediately, keep them in the fridge until analysis.
- c. Analysis of samples is optimally done on the day of collection. When not possible the plasma samples can be stored at 2-8°C. If samples require storage for > 48 hours,

freeze plasma at -20 °C in single aliquots. All samples are allowed to come to room temperature prior to analysis.

- d. Use plasma samples undiluted or diluted to a ratio of 1:2 with deionised water if the volume is insufficient.
- e. Plasma samples that were frozen or stored in the fridge should be vortexed briefly and centrifuged again at ~5000 x g for 2-3 minutes immediately prior to analysis. If necessary, remove fibrin clots using a wooden applicator.

### Analysis:

Samples that produce results that lie outside the linear range for a specific assay have to be re-tested. In some cases it may be necessary to dilute samples with water to bring test results into range.

## Notes

Blood collection for Clinical Chemistry and Hematology is usually performed as a non-fasting, terminal procedure but can be performed as a non-terminal procedure under certain circumstances. Mice from the terminal procedure may be used for subsequent gross pathology and other procedures included in terminal assessments. Whole blood (for Hematology) and plasma (for Clinical Chemistry) require different collection tubes so two independent samples are required from each mouse.

The information about the date of the experiment, that is the date when the measurement is performed, is an important parameter which is to be submitted in the Experiment xml file (dateOfExperiment="2013-02-28").

**Dilution.** Dilution of blood is highly discouraged, but is allowed when the total necessary amount is not obtained. If dilution is necessary then the assays should be done in one run.

**Hemolysis.** Two fields currently exist to capture metadata information about the hemolysis status in the clinical chemistry plasma samples. The first is the LIH Hemolysis severity score which can only be performed by clinics who run one of the Beckman Coulter AU-series of analysers. Such clinics are encouraged to capture and submit the hemolysis score of the LIH test in this field. Clinics who do not have an AU analyser are encouraged to use the second /alternative field which is simply titled Hemolysis. Simply enter "slight", "moderate", or "marked" based on whether the sample is visibly haemolysed or not. Provision of this information is not compulsory and it is suggested that any clinic completes at least one field or the other (not both).

## Data QC

1. Plasma samples must be free of Fibrin clots in order to be analysed.
2. Badly haemolysed samples should be discarded.
3. Each morning, all parameters are tested with control sera (see ESLIM\_015\_001\_Annex\_3: Controls for biochemistry on AU400). Some parameters are tested with control serum level 1 (Beckman Coulter System Reagent, ODC0003)

and control serum level 2 (Beckman Coulter System Reagent, ODC0004), which consists of lyophilised human plasma with a normal and a pathological concentration. Other parameters are tested with specific controls from other suppliers.

4. Controls are thawed and vortexed before utilisation and loaded according to the analyser's display. Control values must lie within the acceptable range indicated by the manufacturer, otherwise the specific tests must be recalibrated and specific measurements repeated. Controls can be stored in 200l aliquots of control sera frozen at -20°C for up to 1 month.

## Metadata and examples

Metadata	Example
Equipment ID	ID of the machine used when more than 1 is used having same model and manufacturer. E.g. machine 1, machine 2, machine Minnie, machine Mickey Mouse, etc.
Equipment manufacturer	Manufacturer of the equipment. E.g. Olympus Diagnostics.
Equipment model	Model of the equipment. E.g. AU400
Blood collection tubes	The tubes used for blood collection. E.g. Sarstedt Li-Heparin gel tubes or Kabe Labortechnik Lithium heparin coated tubes.
Anaesthesia used for blood collection	The drug used for anaesthesia during blood collection. E.g. Isoflurane.
Method of blood collection	Concise description of the method used for blood collection. E.g. retro-orbital puncture.
Anticoagulant	Anticoagulant drug used for blood collection. E.g. Li-Heparin.
Samples kept on ice between collection and analysis	Yes/No.

Storage temperature from blood collection till measurement	E.g. 2°C
Sample status	Indicate if the sample were frozen (analysis on the same day of collection not possible) or fresh (analysis on the same day of collection). E.g Fresh/Frozen.
Plasma dilution	Dilution is highly discouraged but if necessary indicate here. E.g. “No dilution” or 1:2. Note that results submitted to DCC are assumed to be already corrected for any dilutions made.
ID of blood collection SOP	ID of the protocol followed for blood collection. Can be a center specific protocol. E.g. ESLIM_024_001.
Date and time of blood collection	Time of day for collection is in the morning, starting no earlier than 07:30. E.g. Year, month, day, time.
Date of measurement	The day of blood analysis. Year, month, day.
Hemolysis status	If no AU analyser score is provided, indicate here the gauged degree of hemolysis. E.g. slight/moderate/marked.
Blood collection experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures indicating the experimenter who collected the blood. E.g. Harw_001, or 1/2/3.
Blood analysis experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures indicating the experimenter who analyzed the blood. E.g. Harw_001, or 1/2/3.
Date equipment last calibrated	Most recent date in which the equipment (or any part of) used in the procedure was subject to a calibration event.
Date and time of sacrifice	The date and time when the mouse is sacrificed.

# Parameters and Metadata

## Lactate dehydrogenase TCPLA\_CBC\_022\_001 | v1.2

simpleParameter

Req. Analysis: false      Req. Upload: false      Is Annotated: true

Unit Measured: U/l

-----

## Alkaline phosphatase TCPLA\_CBC\_014\_001 | v1.2

simpleParameter

Req. Analysis: false      Req. Upload: true      Is Annotated: true

Unit Measured: U/l

-----

## Calcium TCPLA\_CBC\_009\_001 | v1.5

simpleParameter

Req. Analysis: false      Req. Upload: true      Is Annotated: true

Unit Measured: mg/dl

-----

## Total protein TCPLA\_CBC\_006\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: g/l

---

## LIH (Hemolysis Severity - available on AU analysers) TCPLA\_CBC\_019\_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

---

## Potassium TCPLA\_CBC\_002\_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mmol/l

---

## Date and time of sacrifice TCPLA\_CBC\_040\_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

---

## Phosphorus TCPLA\_CBC\_010\_001 | v1.6

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

---

## Date equipment last calibrated TCPLA\_CBC\_050\_001 | v1.2

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

---

## Date and time of blood collection TCPLA\_CBC\_046\_001 | v1.2

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

---

## Equipment ID TCPLA\_CBC\_033\_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

---

## Glycerol TCPLA\_CBC\_027\_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mmol/l

---

## Triglycerides TCPLA\_CBC\_017\_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

---

## Iron TCPLA\_CBC\_011\_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mg/dl

---



**C-reactive protein** TCPLA\_CBC\_032\_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mg/l

---

**Transferrin** TCPLA\_CBC\_031\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mg/dl

---

**Sample dilution** TCPLA\_CBC\_044\_001 | v1.2

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

**Options:** Neat serum, 1:4, Varies, Yes (by Equipment, automatically), 1:2, Neat plasma, 1:3, 1:5,

---

**Blood collection tubes** TCPLA\_CBC\_039\_001 | v1.1

procedureMetadata

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** false

**Options:** BD Microtainer Lithium Heparin Tube,  
BD Microtainer Lithium Heparin/PST Gel Blood Tube,  
TERUMO CAPIJECT Lithium heparin coated tubes, Eppendorf 1.7ml,  
Kabe Labortechnik Lithium heparin coated tubes, Kabe Labortechnik 1000ul Lithium Heparin,  
Sarstedt Li-Heparin gel tubes, Greiner MiniCollect Lithium Heparin 1ml,

---

## Blood analysis experimenter ID TCPLA\_CBC\_051\_001 | v1.0

procedureMetadata

**Req. Analysis:** false

**Req. Upload:** true

**Is Annotated:** false

## Equipment manufacturer TCPLA\_CBC\_034\_001 | v1.0

procedureMetadata

**Req. Analysis:** true

**Req. Upload:** true

**Is Annotated:** false

**Options:** Cobas, Beckman Coulter, Roche, JEOL (Siemens), Olympus Diagnostics, Hitachi,

---

## Storage temperature from blood collection till measurement TCPLA\_CBC\_041\_001 | v1.3

procedureMetadata

**Req. Analysis:** true

**Req. Upload:** true

**Is Annotated:** false

**Unit Measured:** C

**Options:** 2, 4, 18-22, -80,

---

**Creatine kinase** TCPLA\_CBC\_028\_001 | v1.2

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** U/l

---

**Sodium** TCPLA\_CBC\_001\_001 | v1.3

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** mmol/l

---

**Fasting** TCPLA\_CBC\_057\_001 | v1.0

procedureMetadata

**Req. Analysis:** true

**Req. Upload:** true

**Is Annotated:** false

**Options:** Sixteen hours before bleeding, Four hours before bleeding, No,

---

## Chloride TCPLA\_CBC\_003\_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mmol/l

---

## Magnesium TCPLA\_CBC\_054\_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mg/dl

---

## Glycosilated hemoglobin A1c (HbA1c) TCPLA\_CBC\_052\_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: %

---

# Equipment model TCPLA\_CBC\_035\_001 | v1.0

procedureMetadata

Req. Analysis: true                      Req. Upload: true                      Is Annotated: false

Options: AU 680, Hitachi 917, Integra 400 Plus, JCA-BM6070, 7020, UniCel 600 Pro, JCA-BM2250 (Advia 2400), AU 400, AU 480, Cobas,

-----

# Free fatty acids TCPLA\_CBC\_026\_001 | v1.4

simpleParameter

Req. Analysis: false                      Req. Upload: false                      Is Annotated: true

Unit Measured: mmol/l

-----

# Lipase TCPLA\_CBC\_021\_001 | v1.1

simpleParameter

Req. Analysis: false                      Req. Upload: false                      Is Annotated: true

Unit Measured: U/l

-----

# Alpha-amylase TCPLA\_CBC\_023\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: U/l

---

**Alanine aminotransferase** TCPLA\_CBC\_013\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: U/l

---

**Thyroxine** TCPLA\_CBC\_053\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: ug/dl

---

**Sample status** TCPLA\_CBC\_043\_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Options: Frozen, Fresh and frozen, Fresh,

---

## Cholesterol ratio TCPLA\_CBC\_058\_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Derivation: div('TCPLA\_CBC\_015\_001', 'TCPLA\_CBC\_016\_001')

---

## Samples kept on ice between collection and analysis TCPLA\_CBC\_042\_001 | v1.1

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: No, Yes,

---

## Uric acid TCPLA\_CBC\_029\_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: umol/l

---

**UIBC (unsaturated iron binding capacity)** TCPLA\_CBC\_024\_001 | v1.0

simpleParameter

Req. Analysis: false      Req. Upload: false      Is Annotated: true

Unit Measured: umol/l

---

**Glucose** TCPLA\_CBC\_018\_001 | v1.5

simpleParameter

Req. Analysis: false      Req. Upload: true      Is Annotated: true

Unit Measured: mg/dl

---

**ID of blood collection SOP** TCPLA\_CBC\_045\_001 | v1.1

procedureMetadata

Req. Analysis: false      Req. Upload: true      Is Annotated: false

Options: sop.inv.019, PHENO\_CBC, RIKENMPP\_004a\_003, sop.inv.063, ESLIM\_024\_001,

---

**Total cholesterol** TCPLA\_CBC\_015\_001 | v1.4

simpleParameter



**Req. Analysis:** false

**Req. Upload:** true

**Is Annotated:** true

**Unit Measured:** mg/dl

---

**Method of blood collection** TCPLA\_CBC\_037\_001 | v1.0

procedureMetadata

**Req. Analysis:** true

**Req. Upload:** true

**Is Annotated:** false

**Options:** Heart puncture, Cardiac puncture, Jugular vein, Retro-orbital puncture, Tail vein,

---

**Blood collection experimenter ID** TCPLA\_CBC\_049\_001 | v1.1

procedureMetadata

**Req. Analysis:** false

**Req. Upload:** true

**Is Annotated:** false

---

**Hemolysis status** TCPLA\_CBC\_048\_001 | v1.1

procedureMetadata

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** false

**Options:** None, Moderate, Marked, Slight,

---

**HDL-cholesterol** TCPLA\_CBC\_016\_001 | v1.4

simpleParameter

Req. Analysis: false      Req. Upload: true      Is Annotated: true

Unit Measured: mg/dl

---

**Sample type** TCPLA\_CBC\_056\_001 | v1.0

procedureMetadata

Req. Analysis: true      Req. Upload: true      Is Annotated: false

Options: Plasma, Serum,

---

**LDL-cholesterol** TCPLA\_CBC\_025\_001 | v1.3

simpleParameter

Req. Analysis: false      Req. Upload: false      Is Annotated: true

Unit Measured: mg/dl

---

**Albumin** TCPLA\_CBC\_007\_001 | v1.2

simpleParameter

**Req. Analysis:** false

**Req. Upload:** true

**Is Annotated:** true

**Unit Measured:** g/l

---

## Fructosamine TCPLA\_CBC\_020\_001 | v1.2

simpleParameter

**Req. Analysis:** false

**Req. Upload:** false

**Is Annotated:** true

**Unit Measured:** umol/l

---

## Reagent manufacturer TCPLA\_CBC\_059\_001 | v1.0

procedureMetadata

**Req. Analysis:** true

**Req. Upload:** false

**Is Annotated:** false

**Options:** Beckman Coulter, Microgenics, Wako and Sekisui,

---

## Aspartate aminotransferase TCPLA\_CBC\_012\_001 | v1.2

simpleParameter

**Req. Analysis:** false

**Req. Upload:** true

**Is Annotated:** true

**Unit Measured:** U/l

---

## Total bilirubin TCPLA\_CBC\_008\_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

---

## Urea (Blood Urea Nitrogen - BUN) TCPLA\_CBC\_004\_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

---

## Anticoagulant TCPLA\_CBC\_038\_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Options: Heparine, No, Sodium Heparin, Lithium Heparin,

---

# Anesthesia used for blood collection TCPLA\_CBC\_036\_001 | v1.0

procedureMetadata

Req. Analysis: true                      Req. Upload: true                      Is Annotated: false

**Options:** Injection narcosis with Ketamine (137mg/kg)/Xylazine (6.6mg/kg),  
Injection narcosis with Ketamine (110mg/kg)/Xylazine (11mg/kg),  
Injection narcosis with Sodium Pentobarbital (Euthatal),  
Injection narcosis with Ketamine (110mg/kg)/Xylazine (11mg/kg)/ Antipamezole (Antisedan, 1mg/kg),  
Injection narcosis with Tribromoethanol (Avertin), Gas anaesthesia with Isoflurane,  
Injection narcosis with Ketamine (100mg/kg)/ Xylazine (10mg/kg )/Antipamezole (Antisedan, 1mg/kg),  
Injection narcosis with Sodium Pentobarbital (Pentobarb, 0.1ml), No,  
Injection narcosis with Ketamine (100mg/kg)/Xylazine (10mg/kg),  
Injection narcosis with Sodium Pentobarbital (Somnopentyl),

---

# Ferritin TCPLA\_CBC\_030\_001 | v1.3

simpleParameter

Req. Analysis: false                      Req. Upload: false                      Is Annotated: true

Unit Measured: ng/ml

---

# Creatinine TCPLA\_CBC\_005\_001 | v1.5

simpleParameter

Req. Analysis: false                      Req. Upload: true                      Is Annotated: true

Unit Measured: mg/dl

**Difficult bleed** TCPLA\_CBC\_055\_001 | v1.0  
procedureMetadata

Req. Analysis: false      Req. Upload: false      Is Annotated: false

Options: Yes, No,