

X-ray BCMLA_XRY_001

Purpose

Construct and analyse digital X-ray images in immobilised mice using a Faxitron X-Ray system or NTB digital X-ray scanner.

Experimental Design

- **Minimum number of animals** : 4M + 4F
- **Age at test**: Week 57

Procedure

Perform steps 3.1 and 3.2 unless the mouse is coming from DEXA analysis

3.1 Anesthetize the mice.

3.2 Monitor the animal carefully until unconsciousness by ensuring the mouse is adequately sedated.

3.3 Mount the anesthetized mouse onto an X-ray permeable plate. Curve the tail along the left side of the mouse. (Note: this will enable identification of the X-ray as a dorsal or ventral image.)

Faxitron X-ray system

3.3.1 Insert the plate into the Faxitron X-ray system.

3.3.2 Take an image using the software that is supplied by the X-ray scanner.

3.3.3 Set the machine to a voltage of 25kV with an integration time of 40ms.

3.3.4 Take an image using the iX-Pect software that is supplied by the X-ray scanner.

NTB digital X-ray scanner

3.4 Remove the mouse once the image is captured. Place the mouse on a heated mat, set at 37°C, in a cage and monitor closely until consciousness is fully regained.

3.5 Analyze the image.

3.6 Or continue with ABR.

Notes

The protocol described herein employs anaesthesia using intraperitoneal (IP) administration; however, anaesthesia using isoflurane is also suitable.

[Digit integrity IMPC_XRY_031_001](#) *Definition:* The state of being whole and undivided.

Parameters and Metadata

Lordosis BCMLA_XRY_058_001 | v1.1

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Options: imageOnly, yes, no, unobservable,

Brachydactyly BCMLA_XRY_030_001 | v1.1

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Options: yes, no, unobservable, imageOnly,

Scapulae BCMLA_XRY_006_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Options: unobservable, abnormal, imageOnly, normal,

Number of ribs right BCMLA_XRY_008_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Lumbar processes BCMLA_XRY_065_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: unobservable, imageOnly, abnormal, normal,

Comment on XRay image BCMLA_XRY_032_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

XRay Images Whole Body Dorso Ventral BCMLA_XRY_034_001 | v1.3

seriesMediaParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Increments: Minimum 1

Syndactylism BCMLA_XRY_029_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: imageOnly, unobservable, Syndactylism, no syndactylism,

Pelvis BCMLA_XRY_012_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: imageOnly, normal, abnormal, unobservable,

Number of caudal vertebrae BCMLA_XRY_017_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Tibia length BCMLA_XRY_033_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mm

Settings time of exposure BCMLA_XRY_039_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: s

Xray equipment manufacturer BCMLA_XRY_037_001 | v1.4

procedureMetadata

Req. Analysis: true

Req. Upload: false

Is Annotated: false

Options: Rigaku Corporation (Micro-CT), Bruker, Faxitron X-Ray Corporation,
Faxitron Bioptics LLC,

Digit integrity BCMLA_XRY_031_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: imageOnly, yes, no, unobservable,

Number of pelvic vertebrae BCMLA_XRY_016_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Maxilla/Pre-maxilla BCMLA_XRY_003_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: imageOnly, abnormal, normal, unobservable,

Sacral processes BCMLA_XRY_066_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: abnormal, imageOnly, unobservable, normal,

Processes on vertebrae BCMLA_XRY_020_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: normal, abnormal, imageOnly, unobservable,

Xray equipment model BCMLA_XRY_038_001 | v1.4

procedureMetadata

Req. Analysis: true

Req. Upload: false

Is Annotated: false

Options: UltraFocus 100, MX-20, MX-20-DC12, R_mCT (Micro-CT), MX-20-DC4, UltraFocus DXA, MS-FX Pro,

Fusion of ribs BCMLA_XRY_011_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: not fused, imageOnly, fused, unobservable,

Fusion of processes BCMLA_XRY_061_001 | v1.1

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Options: no, unobservable, yes, imageOnly,

Joints BCMLA_XRY_027_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Options: normal, imageOnly, abnormal, unobservable,

Pelvic vertebrae morphology BCMLA_XRY_073_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Options: imageOnly, unobservable, abnormal, normal,

Experimenter ID (analysis) BCMLA_XRY_042_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Alive BCMLA_XRY_035_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Skull shape BCMLA_XRY_001_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: unobservable, normal, abnormal, imageOnly,

Number of lumbar vertebrae BCMLA_XRY_015_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

XRay Images Skull Lateral Orientation BCMLA_XRY_050_001 | v1.

2

seriesMediaParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Increments: Minimum 1

Tibia BCMLA_XRY_025_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: normal, imageOnly, abnormal, unobservable,

Fibula BCMLA_XRY_026_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: imageOnly, unobservable, normal, abnormal,

Ulna BCMLA_XRY_023_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: unobservable, imageOnly, normal, abnormal,

Femur BCMLA_XRY_024_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: unobservable, abnormal, normal, imageOnly,

Caudal vertebrae morphology BCMLA_XRY_074_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: imageOnly, unobservable, abnormal, normal,

Scoliosis BCMLA_XRY_056_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: unobservable, yes, no, imageOnly,

Scanner equipment model BCMLA_XRY_046_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Date Xray equipment last calibrated BCMLA_XRY_047_001 | v1.3

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Number of ribs left BCMLA_XRY_009_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Shape of vertebrae BCMLA_XRY_018_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Options: normal, imageOnly, unobservable, abnormal,

Teeth BCMLA_XRY_005_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: normal, unobservable, imageOnly, abnormal,

Thoracic vertebrae morphology BCMLA_XRY_071_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: normal, imageOnly, abnormal, unobservable,

Caudal processes BCMLA_XRY_067_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: unobservable, normal, abnormal, imageOnly,

Radius BCMLA_XRY_022_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: imageOnly, abnormal, unobservable, normal,

XRay Images Whole Body Lateral Orientation BCMLA_XRY_048_001 | v1.2

seriesMediaParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Increments: Minimum 1

Scanner equipment ID BCMLA_XRY_044_001 | v1.1

procedureMetadata

Req. Analysis: false **Req. Upload:** false **Is Annotated:** false

Xray equipment ID BCMLA_XRY_036_001 | v1.1

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Voltage settings BCMLA_XRY_043_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

XRay Images Hind Leg and Hip BCMLA_XRY_052_001 | v1.1

seriesMediaParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Increments: Minimum 1

Lumbar vertebrae morphology BCMLA_XRY_072_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: abnormal, imageOnly, unobservable, normal,

Free Ontology BCMLA_XRY_054_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Cervical vertebrae morphology BCMLA_XRY_070_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: imageOnly, normal, abnormal, unobservable,

Shape of ribs BCMLA_XRY_010_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: unobservable, abnormal, imageOnly, normal,

Clavicle BCMLA_XRY_007_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: normal, imageOnly, abnormal, unobservable,

Number of thoracic vertebrae BCMLA_XRY_014_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Fusion of vertebrae BCMLA_XRY_019_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: imageOnly, yes, no, unobservable,

Kyphosis BCMLA_XRY_057_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: imageOnly, no, unobservable, yes,

Rib morphology BCMLA_XRY_069_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: unobservable, abnormal, normal, imageOnly,

Settings level BCMLA_XRY_040_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Number of cervical vertebrae BCMLA_XRY_013_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

XRay Images Forepaw BCMLA_XRY_049_001 | v1.1

seriesMediaParameter

Req. Analysis: false Req. Upload: false Is Annotated: false

Increments: Minimum 1

Shape of ribcage BCMLA_XRY_059_001 | v1.1

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Options: normal, unobservable, imageOnly, abnormal,

Mandibles BCMLA_XRY_004_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Options: normal, abnormal, imageOnly, unobservable,

Experimenter ID (scan) BCMLA_XRY_041_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Scanner equipment manufacturer BCMLA_XRY_045_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Transitional vertebrae BCMLA_XRY_060_001 | v1.1

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Options: unobservable, yes, imageOnly, no,

Shape of spine BCMLA_XRY_055_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: abnormal, imageOnly, normal, unobservable,

Number of digits BCMLA_XRY_028_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Humerus BCMLA_XRY_021_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: abnormal, normal, imageOnly, unobservable,

Date Scanner equipment last calibrated BCMLA_XRY_053_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Thoracic processes BCMLA_XRY_064_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: abnormal, normal, unobservable, imageOnly,

Cervical processes BCMLA_XRY_063_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: unobservable, normal, imageOnly, abnormal,

XRay Images Skull Dorso Ventral Orientation BCMLA_XRY_051_001 | v1.2

seriesMediaParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Increments: Minimum 1

Polysyndactylism BCMLA_XRY_062_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Options: imageOnly, no, yes, unobservable,

Zygomatic bone BCMLA_XRY_002_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Options: abnormal, imageOnly, normal, unobservable,

Missing cranial rib BCMLA_XRY_068_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Options: No, Yes,
