Clinical Chemistry HRWLLA_CBC_003

Purpose

Clinical chemistry determines biochemical parameters in plasma including enzymatic activity, specific substrates and electrolytes.

Ontological description: MP:0001545 – blood physiology abnormalities.

Experimental Design

• Minimum number of animals: 7M + 7F

• Age at test: Week 59

• Sex: We would expect the results of this test to show sexual dimorphism

Equipment

- 1. Clinical chemistry analyser
- 2. Vortex
- 3. Refrigerated centrifuge
- 4. Eppendorf tubes
- 5. Pipettes (200-1000 ul)

Procedure

Set up the clinical chemistry analyser and perform QC analyses of the control reagents in accordance with the equipment guidelines.

Sample collection and preparation:

- a. Collect the appropriate volume of blood required (160-200l of plasma), for the clinical chemistry analyser being used for assessment, preferably using a tube with an anticoagulant with the relevant blood collection procedure (see IMPC protocol Blood collection by retro-orbital puncture). Time of day for collection is in the morning, starting no earlier than 07:30.
- b. Keep whole blood samples on ice until centrifugation. Centrifuge for 10 minutes at 5000 x g in a refrigerated centrifuge set at 8°C. If plasma samples cannot be analysed immediately, keep them in the fridge until analysis.
- c. Analysis of samples is optimally done on the day of collection. When not possible the plasma samples can be stored at 2-8°C. If samples require storage for > 48 hours,

- freeze plasma at -20 °C in single aliquots. All samples are allowed to come to room temperature prior to analysis.
- d. Use plasma samples undiluted or diluted to a ratio of 1:2 with deionised water if the volume is insufficient.
- e. Plasma samples that were frozen or stored in the fridge should be vortexed briefly and centrifuged again at ~5000 x g for 2-3 minutes immediately prior to analysis. If necessary, remove fibrin clots using a wooden applicator.

Analysis:

Samples that produce results that lie outside the linear range for a specific assay have to be re-tested. In some cases it may be necessary to dilute samples with water to bring test results into range.

Notes

Blood collection for Clinical Chemistry and Hematology is usually performed as a non-fasting, terminal procedure but can be performed as a non-terminal procedure under certain circumstances. Mice from the terminal procedure may be used for subsequent gross pathology and other procedures included in terminal assessments. Whole blood (for Hematology) and plasma (for Clinical Chemistry) require different collection tubes so two independent samples are required from each mouse.

The information about the date of the experiment, that is the date when the measurement is performed, is an important parameter which is to be submitted in the Experiment xml file (dateOfExperiment="2013-02-28").

Dilution. Dilution of blood is highly discouraged, but is allowed when the total necessary amount is not obtained. If dilution is necessary then the assays should be done in one run.

Hemolysis. Two fields currently exist to capture metadata information about the hemolysis status in the clinical chemistry plasma samples. The first is the LIH Hemolysis severity score which can only be performed by clinics who run one of the Beckman Coulter AU-series of analysers. Such clinics are encouraged to capture and submit the hemolysis score of the LIH test in this field. Clinics who do not have an AU analyser are encouraged to use the second /alternative field which is simply titled Hemolysis. Simply enter "slight", "moderate", or "marked" based on whether the sample is visibly haemolysed or not. Provision of this information is not compulsory and it is suggested that any clinic completes at least one field or the other (not both).

Data QC

- 1. Plasma samples must be free of Fibrin clots in order to be analysed.
- 2. Badly haemolysed samples should be discarded.
- 3. Each morning, all parameters are tested with control sera (see ESLIM_015_001_Annex_3: Controls for biochemistry on AU400). Some parameters are tested with control serum level 1 (Beckman Coulter System Reagent, ODC0003)

- and control serum level 2 (Beckman Coulter System Reagent, ODC0004), which consists of lyophilised human plasma with a normal and a pathological concentration. Other parameters are tested with specific controls from other suppliers.
- 4. Controls are thawed and vortexed before utilisation and loaded according to the analyser's display. Control values must lie within the acceptable range indicated by the manufacturer, otherwise the specific tests must be recalibrated and specific measurements repeated. Controls can be stored in 200l aliquots of control sera frozen at -20°C for up to 1 month.

Metadata and examples

Metadata	Example
Equipment ID	ID of the machine used when more than 1 is used having same model and manufacturer. E. g. machine 1, machine 2, machine Minnie, machine Mickey Mouse, etc.
Equipment manufacturer	Manufacturer of the equipment. E.g. Olympus Diagnostics.
Equipment model	Model of the equipment. E.g. AU400
Blood collection tubes	The tubes used for blood collection. E.g. Sarstedt Li-Heparin gel tubes or Kabe Labortechnik Lithium heparin coated tubes.
Anaesthesia used for blood collection	The drug used for anaesthesia during blood collection. E. g. Isofluorane.
Method of blood collection	Concise description of the method used for blood collection. E.g. retro-orbital puncture.
Anticoagulant	Anticoagulant drug used for blood collection. E. g. Li-Heparin.
Samples kept on ice between collection and analysis	Yes/No.

Storage temperature from blood collection till measurement	E.g. 2°C
Sample status	Indicate if the sample were frozen (analysis on the same day of collection not possible) or fresh (analysis on the same day of collection). E.g Fresh/Frozen.
Plasma dilution	Dilution is highly discouraged but if necessary indicate here. E.g. "No dilution" or 1:2. Note that results submitted to DCC are assumed to be already corrected for any dilutions made.
ID of blood collection SOP	ID of the protocol followed for blood collection. Can be a center specific protocol. E.g. ESLIM_024_001.
Date and time of blood collection	Time of day for collection is in the morning, starting no earlier than 07:30. E.g. Year, month, day, time.
Date of measurement	The day of blood analysis. Year, month, day.
Hemolysis status	If no AU analyser score is provided, indicate here the gauged degree of hemolysis. E.g. slight/moderate/marked.
Blood collection experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures indicating the experimenter who collected the blood. E.g. Harw_001, or 1/2/3.
Blood analysis experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures indicating the experimenter who analyzed the blood. E.g. Harw_001, or 1/2/3.
Date equipment last calibrated	Most recent date in which the equipment (or any part of) used in the procedure was subject to a calibration event.
Date and time of sacrifice	The date and time when the mouse is sacrified.

Parameters and Metadata

Phosphorus HRWLLA_CBC_010_001 | v1.6

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true
Unit Measured: mg/dl

Anticoagulant HRWLLA_CBC_038_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Options: Heparine, Sodium Heparin, No, Lithium Heparin,

Total cholesterol HRWLLA_CBC_015_001 | v1.4

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Unit Measured: mg/dl

Creatinine HRWLLA_CBC_005_001 | v1.5

simpleParameter

Unit Measured: mmol/l

Req. Analysis: false	Req. Upload: true	Is Annotated: true
Unit Measured: mg/dl		
Blood analysis exprocedureMetadata	perimenter ID HRWLLA	A_CBC_051_001 v1.0
Req. Analysis: false	Req. Upload: true	Is Annotated: false
Blood collection experimenter ID HRWLLA_CBC_049_001 v1.1 procedureMetadata		
Req. Analysis: false	Req. Upload: true	Is Annotated: false
Chloride HRWLLA_CBC_003_001 v1.4 simpleParameter		
Req. Analysis: false	Req. Upload: false	Is Annotated: true

Alpha-amylase HRWLLA_CBC_023_001 | v1.2 simpleParameter Req. Analysis: false Req. Upload: false Is Annotated: true Unit Measured: U/I Hemolysis status HRWLLA_CBC_048_001 | v1.1 procedureMetadata Req. Analysis: false Req. Upload: false Is Annotated: false Options: None, Slight, Moderate, Marked, Difficult bleed HRWLLA_CBC_055_001 | v1.0 procedureMetadata Req. Analysis: false Req. Upload: false Is Annotated: false Options: Yes, No,

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: U/I

Glycosilated hemoglobin A1c (HbA1c) HRWLLA_CBC_052_001 | v1

.3

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: %

.....

Calcium HRWLLA_CBC_009_001 | v1.5

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Unit Measured: mg/dl

Date and time of blood collection HRWLLA_CBC_046_001 | v1.2

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Magnesium HRWLLA_simpleParameter	CBC_054_001 v1.5	
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: mg/dl		
Glucose HRWLLA_CBC simpleParameter	_018_001 v1.5	
Req. Analysis: false	Req. Upload: true	Is Annotated: true
Unit Measured: mg/dl		
Glycerol HRWLLA_CBC simpleParameter	S_027_001 v1.4	
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: mmol/l		

Alkaline phosphatase HRWLLA_CBC_014_001 | v1.2

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true Unit Measured: U/I Reagent manufacturer HRWLLA_CBC_059_001 | v1.0 procedureMetadata Req. Analysis: true Req. Upload: false Is Annotated: false Options: Microgenics, Wako and Sekisui, Beckman Coulter, Thyroxine HRWLLA_CBC_053_001 | v1.2 simpleParameter Req. Analysis: false Req. Upload: false Is Annotated: true Unit Measured: ug/dl

Anesthesia used for blood collection HRWLLA_CBC_036_001 | v1.0

procedureMetadata

Req. Analysis: true Req. Upload: true Is Annotated: false

Options:

Injection narcosis with Ketamine (110mg/kg)/Xylazine (11mg/kg)/ Antipamezole (Antisedan, 1mg/kg),

Injection narcosis with Ketamine (110mg/kg)/Xylazine (11mg/kg),

Injection narcosis with Ketamine (137mg/kg)/Xylazine (6.6mg/kg),

Injection narcosis with Sodium Pentobarbital (Euthatal),

Injection narcosis with Sodium Pentobarbital (Somnopentyl), No,

Gas anaesthesia with Isofluorane,

Injection narcosis with Ketamine (100mg/kg)/Xylazine (10mg/kg),

Injection narcosis with Sodium Pentobarbital (Pentobarb, 0.1ml),

Injection narcosis with Tribromoethanol (Avertin),

Injection narcosis with Ketamine (100mg/kg)/ Xylazine (10mg/kg)/Antipamezole (Antisedan, 1mg/kg),

Topical local anaesthesia with EMLA,

Ferritin HRWLLA_CBC_030_001 | v1.3

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: ng/ml

Storage temperature from blood collection till measurement HRWLLA_CBC_041_001 | v1.3

procedureMetadata

Reg. Analysis: true Reg. Upload: true Is Annotated: false

Unit Measured: C

Options: 18-22, 4, -80, 2,		
Fructosamine HRW simpleParameter	LLA_CBC_020_001 v1.2	
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: umol/l		
Albumin HRWLLA_CB simpleParameter	C_007_001 v1.2	
Req. Analysis: false	Req. Upload: true	Is Annotated: true
Unit Measured: g/l		
Triglycerides HRWLLA_CBC_017_001 v1.4 simpleParameter		
Req. Analysis: false	Req. Upload: true	Is Annotated: true
Unit Measured: mg/dl		

Cholesterol ratio HRWLLA_CBC_058_001 | v1.0

simpleParameter

Reg. Analysis: false Reg. Upload: false Is Annotated: true **Derivation:** div('HRWLLA CBC 015 001', 'HRWLLA CBC 016 001') Sample type HRWLLA_CBC_056_001 | v1.0 procedureMetadata Req. Analysis: true Req. Upload: true Is Annotated: false Options: Plasma, Serum, Equipment ID HRWLLA_CBC_033_001 | v1.0 procedureMetadata Req. Analysis: false Req. Upload: true Is Annotated: false

Date equipment last calibrated HRWLLA_CBC_050_001 | v1.2

procedureMetadata

Reg. Analysis: false Reg. Upload: false Is Annotated: false

C-reactive protein simpleParameter	HRWLLA_CBC_032_001 v ²	1.0
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: mg/l		
Samples kept on ic	ce between collection	on and analysis HRWL
LA_CBC_042_001 v1.1 procedureMetadata		on and analysis in we
Req. Analysis: true	Req. Upload: true	Is Annotated: false
Options: Yes, No,		
Transferrin HRWLLA_simpleParameter	_CBC_031_001 v1.2	
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: mg/dl		

Lipase HRWLLA_CBC_021_001 | v1.1

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true
Unit Measured: U/I

Sample dilution HRWLLA_CBC_044_001 | v1.2

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Options: 1:3, 1:5, Neat serum, 1:2, Yes (by Equipment, automatically), 1:4, Neat plasma,

Varies,

.....

HDL-cholesterol HRWLLA_CBC_016_001 | v1.4

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Unit Measured: mg/dl

Uric acid HRWLLA CBC 029 001 | v1.2

simpleParameter

Reg. Analysis: false Reg. Upload: false Is Annotated: true Unit Measured: umol/l Total protein HRWLLA_CBC_006_001 | v1.2 simpleParameter Req. Analysis: false Req. Upload: true Is Annotated: true Unit Measured: g/l Fasting HRWLLA_CBC_057_001 | v1.0 procedureMetadata Req. Analysis: true Req. Upload: true Is Annotated: false Options: Four hours before bleeding, No, Sixteen hours before bleeding,

Sample status HRWLLA_CBC_043_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Options: Fresh and frozen, Frozen, Fresh,

UIBC (unsaturated iron binding capacity) HRWLLA_CBC_024_001 | v1.0 simpleParameter Req. Analysis: false Req. Upload: false Is Annotated: true Unit Measured: umol/l **Iron** HRWLLA_CBC_011_001 | v1.5 simpleParameter Req. Analysis: false Req. Upload: false Is Annotated: true Unit Measured: mg/dl Lactate dehydrogenase HRWLLA_CBC_022_001 | v1.2 simpleParameter Req. Analysis: false Req. Upload: false Is Annotated: true Unit Measured: U/I

Alanine aminotransferase HRWLLA_CBC_013_001 | v1.2

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true Unit Measured: U/I Sodium HRWLLA CBC 001 001 | v1.3 simpleParameter Req. Analysis: false Req. Upload: false Is Annotated: true Unit Measured: mmol/l Potassium HRWLLA_CBC_002_001 | v1.3 simpleParameter Req. Analysis: false Req. Upload: false Is Annotated: true Unit Measured: mmol/l

LDL-cholesterol HRWLLA_CBC_025_001 | v1.3

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true Unit Measured: mg/dl Aspartate aminotransferase HRWLLA_CBC_012_001 | v1.2 simpleParameter Req. Analysis: false Req. Upload: true Is Annotated: true Unit Measured: U/L Equipment manufacturer HRWLLA_CBC_034_001 | v1.0 procedureMetadata Req. Analysis: true Req. Upload: true Is Annotated: false Options: Roche, Beckman Coulter, JEOL (Siemens), Olympus Diagnostics, Cobas, Hitachi, Method of blood collection HRWLLA CBC 037 001 Lv1.0

procedureMetadata

Req. Analysis: true Req. Upload: true Is Annotated: false

Options: Tail vein, Cardiac puncture, Jugular vein, Heart puncture, Retro-orbital puncture,

ID of blood collection SOP HRWLLA_CBC_045_001 v1.1 procedureMetadata		
Req. Analysis: false	Req. Upload: true	Is Annotated: false
Options: RIKENMPP_004a_	_003, sop.inv.019, ESLIM_024_	001, sop.inv.063, PHENO_CBC,
Equipment model procedureMetadata	HRWLLA_CBC_035_001 v	1.0
Req. Analysis: true	Req. Upload: true	Is Annotated: false
Options: AU 680, Integra 400 Plus, JCA-BM6070, UniCel 600 Pro, Hitachi 917, AU 400, 7020, AU 480, JCA-BM2250 (Advia 2400), DxC AU 700, Cobas,		
Date and time of sacrifice HRWLLA_CBC_040_001 v1.1 procedureMetadata		
Req. Analysis: false	Req. Upload: true	Is Annotated: false

Urea (Blood Urea Nitrogen - BUN) HRWLLA_CBC_004_001 | v1.5

Req. Analysis: false Req. Upload: true Is Annotated: true

Unit Measured: mg/dl

Total bilirubin HRWLLA_CBC_008_001 | v1.4

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Unit Measured: mg/dl

Blood collection tubes HRWLLA_CBC_039_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Options: Eppendorf 1.7ml, Sarstedt Li-Heparin gel tubes,

Kabe Labortechnik Lithium heparin coated tubes,

TERUMO CAPIJECT Lithium heparin coated tubes,

BD Microtainer Lithium Heparin/PST Gel Blood Tube, BD Microtainer Lithium Heparin Tube,

Kabe Labortechnik 1000ul Lithium Heparin, Greiner MiniCollect Lithium Heparin 1ml,

Free fatty acids HRWLLA_CBC_026_001 | v1.4

Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: mmol/l		
LIH (Hemolysis Se	verity - available on	AU analysers) HRWL
LA_CBC_019_001 v1.3 simpleParameter		
Req. Analysis: false	Req. Upload: false	Is Annotated: false